

# CO2 leak test progress

Jul 2<sup>nd</sup> 2015

Yan Ke

- Phone wire splitter will be arrived by next Monday.
- We have 3 working chambers now. Chamber #3 still have 2.18ppm/hour( $1.6 \times 10^{-5} \text{ ccm}/\text{min}$ ) leak rate.
- Data been taken this week on st00132 and st00134 is shown on the other slide
- Short straw been tested:
  - #2: -0.48ppm/hour
  - #4: -0.18ppm/hour
  - #5:-0.28ppm/hour
- Voltage problem
- Reading failure